
## JANUARY 2022 MONTHLY HBOT WEBINAR

#### NFPA Guidelines & SerenaGroup Quality Metrics & Measures

PRESENTED BY THE CLEVELAND CLINIC | AKRON GENERAL





# NFPA 14.2.5.2, NFPA 14.3.1.5.2, NFPA 14.3.1.5.4 & 4.1

- Class B chambers (monoplace) must be able to depressurize from 3 ATA to ambient pressure in less than 2 minutes.
  - This is confirmed in weekly safety chamber checks.
  - As well as in our annual timed fire drills.
  - All personnel is to be trained in all emergency procedures.





## NFPA 14.2.8.2 - 2.1

- All units housing Hyperbaric Chamber are required to have a means of communication to notify the fire department.
- If the building housing the hyperbaric unit has a fire alarm system, the communication system must be a pull – station connected to the system.





#### NFPA 14.2.11.2.5

• The point of exhaust shall be labeled as an Oxygen exhaust.



Labeled "no open flame"



and "no smoking"







#### NFPA 14.3.1.3.2-2.3

- Each facility should have a designated on-site Hyperbaric safety Director that over sees equipment and requirements.
  - Participate in developing procedure.
  - Provide maintenance.
  - Participate in policy development.





#### NFPA 14.3.1.4.6

- When a chamber is occupied, the operator is to be physically present at all times.
  - Maintaining visual access
  - Maintaining audible contact







## NFPA 14.3.1.6.1.3, NFPA 14.3.1.6.4.1 & 4.2, NFPA 14.3.1.6.5 & 5.2, NFPA 14.2.8.1

- Pretreatment checklist should be completed prior to ALL treatments.
- Have a displayed poster of prohibited items that are not to be introduced into the chamber.
  - Cotton garments provided by HBO tech only, are permitted.
  - Patient educated for the duration of treatments of things not allowed in chamber.
  - Patients subjected to pat downs.
  - No hair products, skin products or deodorants allowed.

"If god didn't give it to you and I didn't give it to you, you can't take it in." ~*Tim Mayhugh*~





- If this is not in your HBO Chamber Room, contact Ally George immediately.
- This is a SG Standard.




#### NFPA 14.3.4.2.1.1


- A log will be kept for all testing.
  - Fire drills
  - Maintenances
  - Chamber cycles
  - Safety checks













#### What is a Class B Chamber?





#### What is a Class B Chamber?

## **A Monoplace Chamber**



### A Safety Director must be appointed.





#### Answer 2

### A Safety Director must be appointed.





A prohibited item list must be visible to patients in the chamber room/patient area.





#### Answer 3

## A prohibited item list must be visible to patients in the chamber room/patient area.





## How often is a center's Quality Metric (QM) Dashboard reported?





## How often is a center's Quality Metric (QM) Dashboard reported?

#### WEEKLY



## What are the goals of Quality Metric (QM) Dashboard? *(list 2)*



## What are the goals of Quality Metric (QM) Dashboard? *(list 2)*

## PTA Complement Minimize No Shows







## SerenaGroup HBOT Contact Information

- Matt Schweyer, Chief Compliance Officer | VP Head of Compliance, Reimbursement & Safety | National Safety Director
- mschweyer@serenagroups.com
- C. 214-315-5109

#### SerenaGroup Education Committee

- Ally George = HBO Educator
- Nick Duquette
- Blair Flinn
- Nancy Trafelet
- Jill Schroder





SerenaGroup HBOT Monthly Show Rate (December 2021)

Centers	Program Director	HBO Show Rate
Cleveland Clinic Akron	Nick	72%
АСМН	Erika	98%
Berkshire	Sean	98%
CHI Health CUMC Bergan	Joe	98%
CHI Health Mercy	Joe	94%
Deborah	Megan	95%
Fairview	Jamie	76%
Henry Ford	Eliece	100%
Jackson	Dean	90%
St. Mary's	Katie	100%
St. Joseph Med Ctr	Christine	83%
Via Christi	Nancy	95%
MH The Woodlands	Andrea	100%
Inspira Health – Elmer	Ally	100%



## SerenaGroup Upcoming HBOT Educational Courses

January 20 – 23, 2022 Jackson Hospital (Dean's Center)



• If you need a Hyperbaric Course on-site, please reach out to the Education Committee ASAP. Dr. Serena is looking at locations and dates for 2022.



#### Next Month's Presenter

**DATE:** February 15, 2022

**PRESENTING:** CHI Health CUMC/Bergan Mercy

**TOPIC:** Identifying Hyperbaric Candidates in the Wound Care Center



